

# Laboratory Evaluation Links Some False-Positive COVID-19 Antigen Test Results Observed in a Field Study to a Specific Lot of Test Strips

Alyssa M. Carter,<sup>a,\*</sup> Alexander Vilorio Winnett,<sup>a,\*</sup> Anna E. Romano,<sup>a</sup> Reid Akana,<sup>a</sup> Natasha Shelby,<sup>a</sup> and Rustem F. Ismagilov<sup>a</sup>

Chemistry and Chemical Engineering, California Institute of Technology, Pasadena, California, USA

During a household-transmission field study using COVID-19 antigen rapid diagnostic tests (Ag-RDT), a common test strip lot was identified among 3 participants with false-positive results. In blinded laboratory evaluation, this lot exhibited a significantly higher false-positive rate than other lots. Because a positive Ag-RDT result often prompts action, reducing lot-specific false positives can maintain confidence and actionability of true-positive Ag-RDT results.

**Keywords.** COVID-19; diagnostics; false positive; faulty lot; lateral flow test; Quidel QuickVue.

Antigen rapid diagnostic tests (Ag-RDTs) are increasingly used for detection of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2). Usage of at-home Ag-RDTs in the United States (US) has increased nearly 4-fold among those with self-reported coronavirus disease 2019 (COVID-19)-like illness between the period of Delta (23 August–11 December 2021) to Omicron (19 December 2021–12 March 2022) variant predominance [1]. Ag-RDTs are also used widely for test-to-enter events and serial screening testing in schools and workplaces; for example, in May 2022 [2], the California Department of Public Health began recommending Ag-RDTs as the primary test for COVID-19 in schools [3].

Ag-RDTs typically have very high specificity; of the 51 Ag-RDTs currently authorized for at-home use in the US as of 2 September 2022, all are required to demonstrate false-

positive rates of  $\leq 2\%$  [4]. However, with widespread use imperfect specificity can result in many false-positive results, and at low prevalence of infection, these false positives can represent a large fraction of or even dominate among all positive results [5].

As part of a COVID-19 household-transmission field study in Southern California initiated in November 2021, participants performed a daily at-home nasal swab Ag-RDT (Quidel QuickVue At-Home OTC COVID-19 Test) and self-collected saliva, anterior nares swab, and oropharyngeal swabs for reverse-transcription quantitative polymerase chain reaction (RT-qPCR) testing [6]. This test was selected for the field study because it was one of the first Ag-RDTs to be granted US Food and Drug Administration Emergency Use Authorization [7] and is widely in use in the US and internationally.

In January 2022, interim analysis of the field study showed a string of 24 Ag-RDT positive results from participants who had corresponding negative results in all 3 specimen types tested by RT-qPCR, causing an elevated clinical false-positive rate (Figure 1A). Further investigation revealed a common Ag-RDT strip lot number (152000) among 3 participants with false-positive results. We then investigated the technical false-positive rate of Ag-RDT test strip lot 152000, and other lots acquired for use in the field study, in a controlled laboratory setting.

## METHODS

### Participant Consent Statement

The Ag-RDT field study [6] was approved by the California Institute of Technology Institutional Review Board under protocol number 20-1026. All adult participants in the study provided written informed consent and all minors provided verbal assent accompanied by written parental permission.

### Laboratory Evaluation of Ag-RDT Test Strips

We created contrived specimens using heat-inactivated SARS-CoV-2 particles (BEI, catalog number NR-52286, lot 70034991) spiked into commercial SARS-CoV-2-negative human nasal fluid (Lee BioSolutions, catalog number 991-13-P, lot 03f4044 and catalog number 991-13-P-PreC, lot 09F3280) at concentrations above and below the inferred limit of detection (LOD) for this assay ( $7 \times 10^6$  copies/mL) [6] and applied them to 2 lots of test strips (152194 and 152532) that did not yield any false-positive results among participants in the field study. Contrived specimens with SARS-CoV-2-negative human nasal fluid alone were also applied to 4 Ag-RDT strip lots (152194, 152532, 000202, as well as 152000, the lot common to participants with observed clinical false-positive

Received 03 November 2022; editorial decision 23 December 2022; accepted 03 January 2023; published online 5 January 2023


\*A. M. C. and A. V. W. contributed equally to this work.

Correspondence: Rustem F. Ismagilov, PhD, Chemistry & Chemical Engineering, California Institute of Technology, 1200 E California Blvd, Pasadena, CA 91125 (rustem.admin@caltech.edu).

### Open Forum Infectious Diseases®

© The Author(s) 2023. Published by Oxford University Press on behalf of Infectious Diseases Society of America. This is an Open Access article distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs licence (<https://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial reproduction and distribution of the work, in any medium, provided the original work is not altered or transformed in any way, and that the work is properly cited. For commercial re-use, please contact journals.permissions@oup.com

<https://doi.org/10.1093/ofid/ofac701>



**Figure 1.** Clinical false-positive rate of antigen rapid diagnostic tests (Ag-RDTs) among participants in a coronavirus disease 2019 (COVID-19) household-transmission field study and subsequent laboratory evaluation of technical false-positive rates among Ag-RDT strip lots. *A*, The biweekly clinical false-positive rate for nasal swab Ag-RDT, defined as a positive Ag-RDT at the same timepoint as negative results by reverse-transcription quantitative polymerase chain reaction (RT-qPCR) in saliva, nasal swabs, and oropharyngeal swab specimens. The proportions displayed below each month represent the number of clinical false-positive results over the total number of false-positive and true-negative Ag-RDT results in the field study during each period. Error bars represent 95% confidence interval (CI). *B*, Laboratory evaluation of the technical false-positive rate for 4 Ag-RDT strip lots was performed using severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2)-negative human nasal fluid (see Methods). The proportion of technical false-positive reads to all reads by readers blinded to experimental conditions is displayed below each lot number. *P* values were obtained using an upper-tailed Fisher exact test. Additional details are provided in [Supplementary Table 1](#).

results). The order of contrived specimens and Ag-RDT strip lots was randomized by the operator.

Contrived specimens (20  $\mu$ L) were pipetted onto the swab that came with each Ag-RDT, and the swab was placed into the Ag-RDT tube containing buffer. Manufacturer instructions were then followed exactly [8], by mixing the swab in the buffer for 1 minute, removing the swab, then placing an Ag-RDT strip in the tube and incubating at room temperature for 10 minutes. The result was then interpreted within 5 minutes by 3 readers blinded to the experimental conditions and test strip lot numbers; each trial with a single test strip therefore resulted in 3 independent reads. Readers were provided with the manufacturer instructions for result interpretation [8] and no additional

guidance. Readers were unable to see the interpretations of other readers.

### Statistical Methods

Clinical false-positive results were defined as positive Ag-RDT results reported by a study participant, at the same timepoint when saliva, nasal swab, and oropharyngeal swab specimens collected by the same participant all resulted negative by high-analytical RT-qPCR testing. The clinical false-positive rate was calculated as the number of clinical false-positive Ag-RDT results over all timepoints with false-positive and true-negative Ag-RDT results, using RT-qPCR as the reference standard. The clinical false-positive rate was binned by 2-week periods for visualization (Figure 1A).

Technical false-positive Ag-RDT results were defined as reads interpreted as positive when contrived specimen containing only SARS-CoV-2-negative nasal fluid was tested. The technical false-positive rate was calculated as the number of technical false-positive reads over all reads originating from specimen containing only SARS-CoV-2-negative nasal fluid. The technical false-positive rate was grouped by Ag-RDT strip lot (Figure 1B).

The 95% confidence interval (CI) of both the clinical and technical false-positive rate was calculated using the method described in the Clinical and Laboratory Standards Institute EP12-A2 document [9]. Statistical testing was performed to assess differences in the clinical false-positive rates between time periods in the field study (Figure 1A), and to compare the technical false-positive rates between Ag-RDT strip lots in the laboratory evaluation (Figure 1B); for all analyses we used the Fisher exact test, implemented in Python 3.8.8.

### RESULTS

A significantly elevated clinical false-positive rate was observed among participants in a field study of a COVID-19 Ag-RDT, compared with what had previously been observed in the study ( $P < .01$ , upper-tailed Fisher exact test; Figure 1A). The elevated false-positive rate prompted the identification of a common Ag-RDT strip lot (152000) among 3 participants with multiple, daily clinical false-positive results. We then sought to evaluate the technical false-positive rate of this lot and other lots acquired for use in the field study, through laboratory evaluation.

To confirm that this Ag-RDT could be performed and produce expected results in a laboratory setting, we created contrived specimens with and without SARS-CoV-2 particles. Contrived specimens were applied to 2 Ag-RDT strip lots that had not yielded clinical false-positive results in the field study. Positive reads were expected when nasal fluid with viral concentrations above the LOD were applied to Ag-RDT strips, and negative reads were expected when viral concentrations were below the inferred LOD, and when only

SARS-CoV-2–negative nasal fluid (without any viral particles) was applied. Contrived specimens with SARS-CoV-2 concentrations between  $1.0 \times 10^7$  and  $1.5 \times 10^7$  copies/mL (above the inferred LOD of the Ag-RDT) were interpreted by readers as positive in 8 of 9 reads (3 independent trials each with 3 reads, 1 from each reader); contrived specimens with viral concentrations between  $2.0 \times 10^6$  and  $4.1 \times 10^6$  copies/mL (below the inferred LOD of the Ag-RDT) were interpreted by readers as negative in all 6 reads (Supplementary Table 1). These results confirmed that the Ag-RDT used in the field study yields expected positive and negative results with contrived specimens in a laboratory setting.

To assess the technical false-positive rate of different lots, SARS-CoV-2–negative human nasal fluid (without the addition of viral particles) was applied to Ag-RDT strips from 4 lots: 152194, 152532, 000202, and 152000 (the lot that produced clinical false positives among 3 different participants) (Figure 1B). No false-positive reads were reported for any trial performed on lots 152194, 152532, or 000202. However, 14 of 18 reads from lot 152000 were interpreted by readers as positive, yielding a technical false-positive rate of 77% (95% CI, 55%–91%); 1 read from this lot was interpreted as invalid. Furthermore, at least 1 reader interpreted a positive result for every trial with a lot 152000 test strip (Supplementary Table 1). The false-positive rate of Ag-RDT strip lot 152000 on laboratory evaluation was significantly higher than the false-positive rate observed for the other 3 test strip lots analyzed ( $P < .01$ , upper-tailed Fisher exact test).

## DISCUSSION

In a field study of a COVID-19 Ag-RDT in Southern California, a specific lot of test strips was found to be common among 3 participants (from 3 different households) with false-positive Ag-RDT results. These participants had negative test results in 3 paired high-analytical-sensitivity RT-qPCR assays (saliva, nasal swab, and oropharyngeal swab) that were collected at the same timepoint. Laboratory evaluation confirmed that when SARS-CoV-2–negative nasal fluid was tested with this specific lot of Ag-RDT strips, readers blinded to randomized test conditions and strip lot numbers consistently interpreted results as positive. The laboratory evaluation supports that this lot was likely yielding false-positive results when in use by participants in our field study.

At-home Ag-RDTs are known to have low clinical sensitivity [6, 10, 11] and are likely to produce false-negative results. The low clinical sensitivity of Ag-RDTs is due to both their low-analytical sensitivity (high limits of detection) and, in the US, their authorized use exclusively with nasal swab specimens, which are not always representative of the patient infection status, especially early in the infection [6, 12–16]. The Centers for Disease Control and Prevention (CDC) has recognized the lack

of clinical sensitivity of Ag-RDTs and in September 2022 updated recommendations to Ag-RDT testing protocols to repeat testing 24–48 hours later [17].

False positives are less frequent. The manufacturer of the Quidel QuickVue At-Home OTC COVID-19 Test, which was not involved in the design or execution of this study, reports a 99.2 negative percent agreement [8], and Ag-RDTs generally have >97% clinical specificity in field evaluations [11, 18]. By late 2020, the CDC recommended a confirmatory nucleic acid amplification test for Ag-RDT–positive results in cases with low pretest probability [19]. However as of April 2022, a single positive result now typically prompts immediate action from individuals, their close contacts, and healthcare personnel [20]. Notably, the Emergency Use Authorization for the Quidel QuickVue At-Home OTC COVID-19 Test [21] encourages individuals who test positive to self-isolate and contact their healthcare provider for follow-up care, which may include additional testing. Therefore, false-positive results can prompt unnecessary isolation and quarantine, needless treatment, consumption of additional testing resources, and diversion of contact tracing efforts from true-positive cases [22]. Further, false-positive results undermine trust in positive Ag-RDT results, such that isolation, treatment, additional testing, and contact tracing may not be initiated when it is appropriate.

False-positive Ag-RDT results are not unique to the current COVID-19 pandemic. The Quidel QuickVue Influenza A + B Test, another Ag-RDT that uses nasal swab specimens, is reported by the manufacturer to have >97 negative percent agreement [23], but during the 2009 influenza A(H1N1) pandemic, the clinical performance of the test resulted in a 62.2 negative percent agreement against RT-PCR [24].

COVID-19 Ag-RDT false-positive results have been reported in a number of contexts [18, 25, 26]. In a recent evaluation of the Quidel QuickVue At-Home OTC COVID-19 test in a college community [27], 8 of 11 participants with positive Ag-RDT results were found to be negative on RT-PCR testing within 24 hours. No definitive cause for these false-positive results was identified.

False-positive results may occur due to a variety of reasons [22, 25, 28–30], including user error, invalid test conditions, improper storage or manufacturing errors that affect reagent chemistry, or off-target binding of human or microbial material (including viruses other than SARS-CoV-2); for example, infection of human rhinovirus A has produced false-positive results in SARS-CoV-2 Ag-RDTs [26]. However, both we and others [31, 32] have found false-positive Ag-RDT results traceable to specific lots. Importantly, the overall false-positive rate observed among participants in our field study was 2.8% (95% CI, 2.1%–3.9%) [6]; monitoring only an overall false-positive rate across lots could mask specific lots with higher false-positive rates.

Lot issues can arise during manufacturing and transportation or can be due to storage conditions after distribution [33, 34]. In our study, Ag-RDTs were stored at room temperature and the mild winter climate in Southern California ensured that temperatures were stable during shipment as well. Here, we demonstrate through a controlled laboratory evaluation that false-positive results captured in a field study of Ag-RDTs were not due to operator error but were lot specific. Therefore, efforts to monitor for lot-dependent false positives (and whether they originate from issues at the manufacturer or distributor/retailer level) can increase the confidence and actionability of positive Ag-RDT results.

### Supplementary Data

**Supplementary materials** are available at *Open Forum Infectious Diseases* online. Consisting of data provided by the authors to benefit the reader, the posted materials are not copyedited and are the sole responsibility of the authors, so questions or comments should be addressed to the corresponding author.

### Notes

**Acknowledgments.** We thank the study participants for making this work possible. We thank Study Coordinators Saharai Caldera for suggesting the laboratory evaluation and both Saharai and Hannah Davich for removing lot 152000 from circulation in the study. We thank Taikun Yamada, John Raymond B. Reyna, Paolo Piatti, and Yap Ching Chew for performing and verifying the validity of RT-qPCR results as previously reported. Finally, we thank all the case investigators and contact tracers at the Pasadena Public Health Department and Caltech Student Wellness Services for their efforts in study recruitment and their work in the pandemic response.

**Data availability.** The data underlying the results presented in the study are available at CaltechDATA at: <https://doi.org/10.22002/fmz6a-0x036>.

**Financial support.** This work was supported by the Ronald and Maxine Linde Center for New Initiatives at the California Institute of Technology and the Jacobs Institute for Molecular Engineering for Medicine at the California Institute of Technology. A. V. W. is supported by a Geffen Fellowship at the David Geffen School of Medicine, University of California, Los Angeles.

**Potential conflicts of interest.** R. F. I. is a co-founder, consultant, and director for and has stock ownership in Talis Biomedical Corporation. All other authors report no potential conflicts.

### References

- Rader B, Gertz A, Luliano AD, et al. Use of at-home COVID-19 tests—United States, August 23, 2021–March 12, 2022. *MMWR Morb Mortal Wkly Rep* **2022**; 71:489–94.
- Aragon T. Preliminary testing framework for K–12 schools for the 2022–2023 school year. **2022**. <https://www.cdph.ca.gov/Programs/CID/DCDC/Pages/COVID-19/Schools-Testing-Framework-K-12-Schools-2022-23.aspx>. Accessed 13 October 2022.
- Aragon T. Testing framework for K–12 schools for the 2022–2023 school year. **2022**. <https://www.cdph.ca.gov/Programs/CID/DCDC/Pages/COVID-19/Testing-Framework-for-K-12-Schools-for-the-2022-2023-School-Year.aspx>. Accessed 25 October 2022.
- US Food and Drug Administration. In vitro diagnostics EUAs: molecular and antigen home use test template. **2021**. <https://www.fda.gov/media/140615/download>. Accessed 13 October 2022.
- Healy B, Khan A, Metezai H, et al. Covid-19 testing, low prevalence and the impact of false positive results. *BMJ* **2020**; 369:m1808.
- Viloria Winnett AV, Akana R, Shelby N, et al. Extreme differences in SARS-CoV-2 Omicron viral loads among specimen types drives poor performance of nasal rapid antigen tests for detecting presumably pre-infectious and infectious individuals, predicting improved performance of combination specimen antigen tests. medRxiv [Preprint]. Posted online 15 July 2022. <https://doi.org/10.1101/2022.07.13.22277513>
- US Food and Drug Administration. In vitro diagnostics EUAs—antigen diagnostic tests for SARS-CoV-2. **2022**. <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas-antigen-diagnostic-tests-sars-cov-2>. Accessed 16 December 2022.
- Quidel. QuickVue at-home OTC COVID-19 test user instructions. <https://www.quidel.com/media/147250/download>. Accessed 13 October 2022.
- Clinical and Laboratory Standards Institute (CLSI). EP12-A2 user protocol for evaluation of qualitative test performance. 2nd ed. Wayne, PA: CLSI; **2008**.
- Calderón RI, Jhaveri TA, Tovar MA, et al. Diagnostic performance assessment of saliva RT-PCR and nasopharyngeal antigen for the detection of SARS-CoV-2 in Peru. *Microbiol Spectr* **2022**; 10:e00861-22.
- Tsao J, Kussman AL, Costales C, Pinsky BA, Abrams GD, Hwang CE. Accuracy of rapid antigen vs reverse transcriptase–polymerase chain reaction testing for SARS-CoV-2 infection in college athletes during prevalence of the Omicron variant. *JAMA Netw Open* **2022**; 5:e2217234.
- Savelle ES, Viloria Winnett AV, Romano AE, et al. Quantitative SARS-CoV-2 viral-load curves in paired saliva samples and nasal swabs inform appropriate respiratory sampling site and analytical test sensitivity required for earliest viral detection. *J Clin Microbiol* **2022**; 60:e0178521.
- Viloria Winnett AV, Akana R, Shelby N, et al. SARS-CoV-2 exhibits extreme differences in early viral loads among specimen types suggesting improved detection of pre-infectious and infectious individuals using combination specimen types. medRxiv [Preprint]. Posted online 15 July 2022. <https://doi.org/10.1101/2022.07.13.22277113>
- Teo AKJ, Choudhury Y, Tan IB, et al. Saliva is more sensitive than nasopharyngeal or nasal swabs for diagnosis of asymptomatic and mild COVID-19 infection. *Sci Rep* **2021**; 11:3134.
- Wyllie AL, Fournier J, Casanovas-Massana A, et al. Saliva or nasopharyngeal swab specimens for detection of SARS-CoV-2. *N Engl J Med* **2020**; 383:1283–6.
- Adamson B, Sikka R, Wyllie AL, Premrurit P. Discordant SARS-CoV-2 PCR and rapid antigen test results when infectious: a December 2021 occupational case series. medRxiv [Preprint]. Posted online 5 January 2022. <https://doi.org/10.1101/2022.01.04.22268770>
- US Centers for Disease Control and Prevention. Self-testing at home or anywhere. **2022**. <https://www.cdc.gov/coronavirus/2019-ncov/testing/self-testing.html>. Accessed 13 October 2022.
- Nelson EJ, McKune SL, Ryan KA, et al. Antigen vs RT-PCR tests for screening quarantined students in Florida during the COVID-19 pandemic SARS-CoV-2 Delta variant surge. *JAMA Pediatr* **2022**; 176:525–6.
- Procop GW, Kadkhoda K, Rhoads DD, Gordon SG, Reddy AJ. Home testing for COVID-19: benefits and limitations [published online ahead of print 1 March 2021]. *Cleve Clin J Med* **2021**. <https://doi.org/10.3949/ccjm.88a.ccc071>
- US Centers for Disease Control and Prevention. Guidance for antigen testing for SARS-CoV-2 for healthcare providers testing individuals in the community. **2022**. Available at: <https://www.cdc.gov/coronavirus/2019-ncov/lab/resources/antigen-tests-guidelines.html>. Accessed 13 December 2022.
- US Food and Drug Administration. QuickVue at-home OTC COVID-19 test—letter of authorization. QuickVue At-Home OTC COVID-19 Test—letter of authorization. **2021**. <https://www.fda.gov/media/147247/download>. Accessed 13 December 2022.
- Brooks ZC, Das S. COVID-19 testing: impact of prevalence, sensitivity, and specificity on patient risk and cost. *Am J Clin Pathol* **2022**; 154:575–84.
- Quidel Corporation. Quidel QuickVue Influenza A + B Test. **2022**. <https://www.quidel.com/sites/default/files/product/documents/EF1350314EN00.pdf>. Accessed 15 December 2022.
- Stevenson HL, Loeffelholz MJ. Poor positive accuracy of QuickVue rapid antigen tests during the influenza A (H1N1) 2009 pandemic. *J Clin Microbiol* **2010**; 48:3729–31.
- Itoh K, Kawamitsu T, Osaka Y, et al. False-positive results in severe acute respiratory coronavirus 2 (SARS-CoV-2) rapid antigen tests for inpatients. *J Infect Chemother* **2021**; 27:1089–91.
- Hase R, Kurita T, Mito H, et al. Potential for false-positive results with quantitative antigen tests for SARS-CoV-2: a case of a child with acute respiratory infection. *J Infect Chemother* **2022**; 28:319–20.
- Tinker SC, Prince-Guerra JL, Vermandere K, et al. Evaluation of self-administered antigen testing in a college setting. *Virol J* **2022**; 19:202.
- US Food and Drug Administration. Potential for false-positive results with antigen tests for rapid detection of SARS-CoV-2—letter to clinical laboratory staff and health care providers. **2020**. <https://www.fda.gov/medical-devices/letters-health-care-providers/potential-false-positive-results-antigen-tests-rapid-detection-sars-cov-2-letter-clinical-laboratory>. Accessed 13 December 2022.

29. Otake S, Miyamoto S, Mori A, Iwamoto T, Kasai M. False-positive results in SARS-CoV-2 antigen test with rhinovirus-A infection. *Pediatr Int* **2021**; 63:1135–7.
30. Patriquin G, Davidson RJ, Hatchette TF, et al. Generation of false-positive SARS-CoV-2 antigen results with testing conditions outside manufacturer recommendations: a scientific approach to pandemic misinformation. *Microbiol Spectr* **2021**; 9:e00683-21.
31. Gans JS, Goldfarb A, Agrawal AK, Sennik S, Stein J, Rosella L. False-positive results in rapid antigen tests for SARS-CoV-2. *JAMA* **2022**; 327:485–6.
32. US Food and Drug Administration. Ellume recalls COVID-19 home test for potential false-positive SARS-CoV-2 test results. **2021**. <https://www.fda.gov/medical-devices/medical-device-recalls/ellume-recalls-covid-19-home-test-potential-false-positive-sars-cov-2-test-results>. Accessed 13 December 2022.
33. Williams C. Did mishandling of Utah rapid antigen tests lead them to be pulled? **2022**. <https://www.ksl.com/article/50344932/did-mishandling-of-utah-rapid-antigen-tests-lead-them-to-be-pulled>. Accessed 13 October 2022.
34. Haage V, Ferreira de Oliveira-Filho E, Moreira-Soto A, et al. Impaired performance of SARS-CoV-2 antigen-detecting rapid diagnostic tests at elevated and low temperatures. *J Clin Virol* **2021**; 138:104796.