

Journal of Biomedical Optics

BiomedicalOptics.SPIEDigitalLibrary.org

Errata: Fourier ptychographic microscopy for filtration-based circulating tumor cell enumeration and analysis

Anthony Williams
Jaebum Chung
Xiaoze Ou
Guoan Zheng
Siddarth Rawal
Zheng Ao
Ram Datar
Changhuei Yang
Richard Cotea

SPIE.

Errata: Fourier ptychographic microscopy for filtration-based circulating tumor cell enumeration and analysis

Anthony Williams,^{a,b} Jaebum Chung,^c Xiaoze Ou,^c Guoan Zheng,^c Siddarth Rawal,^{a,b} Zheng Ao,^{a,b} Ram Datar,^{a,b} Changhuei Yang,^c and Richard Cotea^b

^aUniversity of Miami, Miller School of Medicine, Department of Pathology, 1501 NW 10th Avenue BRB 742, Miami, Florida 33136

^bUniversity of Miami, Dr. John T. Macdonald Foundation Biomedical Nanotechnology Institute (BioNIUM), 1501 NW 10th Avenue BRB 714, Miami Florida 33136

^cCalifornia Institute of Technology, Departments of Electrical Engineering, Bioengineering, and Medical Engineering, 1200 East California Boulevard MC 136-93, Pasadena, California 91125

[DOI: [10.1117/1.JBO.20.12.129802](https://doi.org/10.1117/1.JBO.20.12.129802)]

This article [*J. Biomed. Opt.* **19**(6), 066007 (2014)] was originally published online on 20 June 2014 with an error in the Discussion section. On p. 7, the left column, line 6, the author has clarified that the specified resolution is half-pitch. The sentence has been corrected as follows:

“Using our current experiment microfilter as an example (desired FOV of 6-mm diameter, half-pitch resolution of

0.78 μm , objective's depth of field of $\sim 5 \mu\text{m}$, and a total sample height variation of 250 μm), the total amount of data that a conventional WSI system would have to collect for a color image would be equal to ~ 8.9 gigapixels, one-third of which is assigned to each of the RGB channels.”

This article was corrected online on 7 December 2015.