

# PROCEEDINGS OF SPIE

SPIEDigitalLibrary.org/conference-proceedings-of-spie

## Molecular photoacoustic imaging using gold nanoparticles as a contrast agent

Chulhong Kim, Eun Chul Cho, Jingyi Chen, Kwang Hyun Song, Leslie Au, et al.

Chulhong Kim, Eun Chul Cho, Jingyi Chen, Kwang Hyun Song, Leslie Au, Christopher P. Favazza, Qiang Zhang, Claire M. Cobley, Younan Xia, Lihong V. Wang, "Molecular photoacoustic imaging using gold nanoparticles as a contrast agent," Proc. SPIE 7564, Photons Plus Ultrasound: Imaging and Sensing 2010, 75641V (25 February 2010); doi: 10.1117/12.843549

**SPIE.**

Event: SPIE BiOS, 2010, San Francisco, California, United States

# Molecular photoacoustic imaging using gold nanoparticles as a contrast agent

Chulhong Kim<sup>1</sup>, Eun Chul Cho<sup>1</sup>, Jingyi Chen<sup>1</sup>, Kwang Hyun Song<sup>2</sup>, Leslie Au<sup>1</sup>, Christopher P. Favazza<sup>1</sup>, Qiang Zhang<sup>1</sup>, Claire M. Cobley<sup>1</sup>, Younan Xia<sup>1</sup>, and Lihong V. Wang<sup>1\*</sup>

<sup>1</sup>Department of Biomedical Engineering, Washington University in St. Louis, Campus Box 1097,  
One Brookings Dr. St. Louis, Missouri, 63130

<sup>2</sup>Division of Medical Physics & Engineering, Department of Radiation Oncology, The University of Texas Southwestern Medical Center at Dallas, 5323 Harry Hines Boulevard, Dallas, Texas 75390

\*Corresponding author: [lhwang@biomed.wustl.edu](mailto:lhwang@biomed.wustl.edu)

## ABSTRACT

Gold nanoparticles have received much attention due to their potential diagnostic and therapeutic applications. Gold nanoparticles are attractive in many biomedical applications because of their biocompatibility, easily modifiable surfaces for targeting, lack of heavy metal toxicity, wide range of sizes (35–100 nm), tunable plasmonic resonance peak, encapsulated site-specific drug delivery, and strong optical absorption in the near-infrared regime. Specifically, due to their strong optical absorption, gold nanoparticles have been used as a contrast agent for molecular photoacoustic (PA) imaging of tumor. The plasmonic resonance peak of the gold nanocages (AuNCs) was tuned to the near-infrared region, and the ratio of the absorption cross-section to the extinction cross-section was approximately ~70%, as measured by PA sensing. We used PEGylated gold nanocages (PEG-AuNCs) as a passive targeting contrast agent on melanomas. After 6-h intravenous injection of PEG-AuNCs, PA amplitude was increased by ~14 %. These results strongly suggest PA imaging paired with AuNCs is a promising diagnostic tool for early cancer detection.

**Keywords:** Photoacoustic tomography, molecular imaging, gold nanoparticles, B16 melanoma.

## INTRODUCTION

Photoacoustic imaging is a hybrid biomedical imaging modality that can provide strong optical absorption contrasts with high ultrasonic spatial resolution.<sup>1,2</sup> Since ultrasonic parameters determine the spatial resolution beyond one optical transport mean free path (~1mm) in PA imaging, the maximum imaging depth and resolution are scalable while diffusive photons are available.

The use of near-infrared (NIR) light is highly desirable to increase the penetration depth in optical imaging because of the weak optical absorption of hemoglobin and scattering of tissues in this spectral region. However, it is a challenge to assess tumors in deep tissues using only morphological and functional PA imaging based on intrinsic contrasts. Therefore, molecularly specific exogenous contrast agents are desired to enhance the sensitivity and specificity on tumor assessment using PA imaging in deep tissues. Organic dyes such as IRDye-800 and indocyanine green have been used as contrast agents for PA imaging,<sup>3</sup> but these small molecules suffer from rapid blood circulation time and relatively small optical absorption cross-sections. Recently, gold nanostructures<sup>4</sup> and single-walled carbon nanotubes<sup>5</sup> have been demonstrated as contrast agents for PA imaging *in vivo*. Practically, the toxicity of the carbon nanotubes is in question for *in-vivo* imaging. Gold nanostructures are promising candidates because of their strong light absorption in the NIR regime, bio-compatibility, and easy functionalization of surfaces with tumor recognition moieties. In this Proceeding, we demonstrate the use of PEG-AuNCs as a passive targeting contrast agent in PA imaging of B16 melanomas *in vivo*.<sup>6,7</sup>

## METHODS AND MATERIALS

A photograph of a dark-field confocal photoacoustic imaging system is shown in Figure 2. A Ti:sapphire laser (LT-2211A, LOTIS TII) pumped by a Q-switched Nd:YAG laser (LS-2137, LOTIS) with 6-ns pulse duration and 10-Hz pulse repetition rate was used. Formed by a spherical conical lens and an optical condenser, the donut-shaped light illumination was coaxially aligned with the ultrasound focus in water. Dark-field confocal configuration provided a greater penetration depth and higher SNR. The light fluences on the skin were less than  $6.9 \text{ mJ/cm}^2$ , within the ANSI limit. A 10-MHz single-element ultrasound transducer was used as a detector. The axial and transverse resolutions are 125  $\mu\text{m}$  and 140  $\mu\text{m}$ , respectively. By measuring PA amplitudes according to their arrival times, one-dimensional depth-resolved image could be acquired, referred to as an A-line. Additional raster scanning along the one transverse direction enables the formation of two-dimensional depth-resolved images (B-scans), and further raster scanning along another transverse direction provides three-dimensional images of optical absorption heterogeneities in biological tissues. The acquired three-dimensional raw data can be processed in two forms: a maximum amplitude projection (MAP), a projection of the maximum PA amplitude along each A-line onto the corresponding plane, and a true three-dimensional image using Volview software (Kitware).




Figure 1: Photograph of volumetric photoacoustic imaging system. Ti:Sa laser; Ti:Sapphire laser, AMP; amplifier, AD; analog digital converter, PD; photodiode, WT; water tank, SH; sample holder.

Nude mice weighing about 30 g were used for the *in-vivo* experiments. All *in vivo* animal experiments were carried out in compliance with the Washington University Institutional Animal Care and Use Committee. The mice were initially anesthetized with a mixture of ketamine (85 mg/kg) and xylazine (15 mg/kg). During the PA imaging experiments, anesthesia was maintained using vaporized isoflurane (1 L/min oxygen and 0.75% isoflurane, Euthanex Corp.), and vitals were monitored using a pulse oximeter (NONIN Medical INC., 8600V). The body temperatures of mice were maintained by a water heating pad.

## RESULTS AND DISCUSSIONS

We investigated the feasibility of PA imaging of B16 melanomas in nude mice using PEG-AuNCs (Fig. 2). Prior to injection, a control image was acquired. Then, 100  $\mu\text{l}$  of PEG-AuNCs was intravenously injected to mice via the

tailvein at concentration of 10 nM. A series of noninvasive volumetric PA images were obtained up to 6 h after injection *in vivo*. We could clearly see the enhancement of PA amplitude as time elapsed (Fig. 2).




Figure 2: *In vivo* noninvasive PA MAP images of a B16 melanoma over time before and after injection of PEG-AuNCs

Figure 3 shows B-scan PA images from the same position of the animal before and after injection. The location of the melanoma was maintained.




Figure 3: B-scan PA images of melanomas before and after injection of PEG-AuNCs.

Figure 4 shows plots for the enhancement of PA signals in the tumors over a period of time. The PA signals with PEG-AuNCs show a monotonous increase until 5 h post-injection and reached a level-off. The PA amplitude increased up to 14% at 6 h post-injection.




Figure 4: Increase in PA amplitudes within melanomas after intravenous injection of PEG-AuNCs.

## CONCLUSIONS

In summary, we have demonstrated that passive targeting PEG-AuNCs can be used as a PA contrast agent on B16 melanomas *in vivo*. The detection sensitivity of the PA imaging system was  $\sim 4.5$  pM. Biocompatible AuNCs can be used as a diagnostic contrast agent for molecular PA imaging, a photothermal agent based on strong optical absorption, and a drug-delivery platform based on a hollow feature. Moreover, the circulation time of PEG-AuNCs in the blood stream is relatively long compared to organic dyes, which is an additional benefit for *in vivo* imaging.<sup>12</sup> Potentially, high-resolution 3D morphological and functional PA imaging, when combined with AuNCs can diagnose tumors at early stages, treat tumors using either the photothermal effect or drug-delivery, monitor post-treatment processes, and further improve post-tumor treatment plans in real clinical application.

## ACKNOWLEDGEMENT

This work was supported in part by grants from National Institutes of Health (R01 EB000712, R01 EB008085, and U54 CA136398 -- the Network for Translational Research -- to L.V.W.). L.V.W. has a financial interest in Microphotoacoustics, Inc. and Endra, Inc., which, however, did not support this work. Y.X. thanks the National Institutes of Health for a 2006 Director's Pioneer Award (DP1 OD000798-04). E.C.C. was also partially supported by a fellowship award from the Korea Research Foundation (KRF-2007-357-D00070) funded by the Korean Government. J.C. was supported by a Pilot Grant from Washington University Molecular Imaging Center.

## REFERENCES

- [1] X. D. Wang, et al., "Noninvasive laser-induced photoacoustic tomography for structural and functional *in vivo* imaging of the brain," *Nat Biotechnol* **21**, 803-806 (2003).
- [2] H. F. Zhang, K. Maslov, G. Stoica, and L. V. Wang, "Functional photoacoustic microscopy for high-resolution and noninvasive *in vivo* imaging," *Nat Biotechnol*. **24(7)**, 848-851 (2006).
- [3] M. L. Li, et al., "Simultaneous molecular and hypoxia imaging of brain tumors *in vivo* using spectroscopic photoacoustic tomography," *P IEEE* **96**, 481-489 (2008).
- [4] P.C. Li, et al., "In *vivo* Photoacoustic Molecular Imaging with Simultaneous Multiple Selective Targeting Using Antibody-Conjugated Gold Nanorods," *Optics Express* **16**, 18605-18615 (2008).

- [5] A. De la Zerda, et al. "Carbon nanotubes as photoacoustic molecular imaging agents in living mice," *Nat Nanotechnol* **3**, 557-562 (2008).
- [6] J. Y. Chen, et al., "Gold nanocages: Engineering their structure for biomedical applications," *Adv Mater* **17**, 2255-2261 (2005).
- [7] S. E. Skrabalak, L. Au, X. D. Li, and Y. Xia, "Facile synthesis of Ag nanocubes and Au nanocages," *Nat Protoc* **2**, 2182-2190 (2007).