

Lock-in camera based heterodyne holography for ultrasound-modulated optical tomography inside dynamic scattering media

Cite as: Appl. Phys. Lett. **108**, 231106 (2016); <https://doi.org/10.1063/1.4953630>

Submitted: 09 February 2016 . Accepted: 28 May 2016 . Published Online: 08 June 2016

Yan Liu, Yuecheng Shen , Cheng Ma, Junhui Shi, and Lihong V. Wang



View Online



Export Citation



CrossMark

ARTICLES YOU MAY BE INTERESTED IN

[High-speed single-shot optical focusing through dynamic scattering media with full-phase wavefront shaping](#)

Applied Physics Letters **111**, 221109 (2017); <https://doi.org/10.1063/1.5009113>

[Focusing light through scattering media by polarization modulation based generalized digital optical phase conjugation](#)

Applied Physics Letters **111**, 201108 (2017); <https://doi.org/10.1063/1.5005831>

[Pulsed ultrasound-modulated optical tomography using spectral-hole burning as a narrowband spectral filter](#)

Applied Physics Letters **93**, 011111 (2008); <https://doi.org/10.1063/1.2952489>



Lock-in camera based heterodyne holography for ultrasound-modulated optical tomography inside dynamic scattering media

Yan Liu, Yuecheng Shen, Cheng Ma, Junhui Shi, and Lihong V. Wang^{a)}

Optical Imaging Laboratory, Department of Biomedical Engineering, Washington University in St. Louis, St. Louis, Missouri 63130, USA

(Received 9 February 2016; accepted 28 May 2016; published online 8 June 2016)

Ultrasound-modulated optical tomography (UOT) images optical contrast deep inside scattering media. Heterodyne holography based UOT is a promising technique that uses a camera for parallel speckle detection. In previous works, the speed of data acquisition was limited by the low frame rates of conventional cameras. In addition, when the signal-to-background ratio was low, these cameras wasted most of their bits representing an informationless background, resulting in extremely low efficiencies in the use of bits. Here, using a lock-in camera, we increase the bit efficiency and reduce the data transfer load by digitizing only the signal after rejecting the background. Moreover, compared with the conventional four-frame based amplitude measurement method, our single-frame method is more immune to speckle decorrelation. Using lock-in camera based UOT with an integration time of 286 μs , we imaged an absorptive object buried inside a dynamic scattering medium exhibiting a speckle correlation time (τ_c) as short as 26 μs . Since our method can tolerate speckle decorrelation faster than that found in living biological tissue ($\tau_c \sim 100\text{--}1000 \mu\text{s}$), it is promising for *in vivo* deep tissue non-invasive imaging. *Published by AIP Publishing.*
[\[http://dx.doi.org/10.1063/1.4953630\]](http://dx.doi.org/10.1063/1.4953630)

Light scattering prevents optical imaging from achieving high resolution deep inside scattering media. To break this optical diffusion limit (~ 1 mm in tissue), ultrasound-modulated optical tomography (UOT),^{1,2} also called acousto-optic imaging,³ has been developed to image optical absorption⁴ and scattering⁵ properties with ultrasound-determined spatial resolution at depth up to several centimeters. Due to the acousto-optic effect, light passing through an ultrasonic beam undergoes a frequency shift to multiples of the ultrasonic frequency.^{6,7} By detecting the frequency-shifted light (tagged light), ultrasonically defined spatial resolution can be reached.

Various methods have been developed to detect the very few ultrasonically tagged photons out of a large background of untagged photons.^{2,3} Initially, a single-pixel detector, such as a photomultiplier tube^{4,7} or a photodiode,⁸ was employed. Since each speckle grain oscillates with a different phase, when N speckle grains fall on the detector, the useful AC signal amplitude is proportional to \sqrt{N} , while the informationless DC background is proportional to N .^{1,8} Thus, the largest modulation depth (defined as the ratio of the AC signal to the DC background) is obtained when the detector detects no more than one speckle grain, which severely limits the detection etendue⁹ (defined as the product of the detection area and the acceptance solid angle). To increase the detection etendue without reducing the modulation depth, three types of methods have been developed. The first type relies on a narrow spectral filter (\sim MHz) to filter out the untagged light, so that a large-area single-pixel detector can be used. Examples include Fabry-Perot interferometers^{10–12} and spectral-hole burning^{13–15} based methods. These techniques

are immune to speckle decorrelation due to motions of scatterers, but require bulky and expensive equipment. The second type of method uses crystal based holography to convert the wavefront of a reference beam to the complex wavefront of the sample beam, so the two beams can interfere constructively on a large single-pixel detector.^{16–19} This method, however, can be affected by speckle decorrelation due to scatterers' motion inside living biological tissue, since the response time of the crystal is usually much longer than the speckle correlation time ($\sim 0.1\text{--}1$ ms) of living tissue.^{20–22} Recently, Ramaz's group demonstrated the promise of $\text{Sn}_2\text{P}_2\text{S}_6:\text{Te}$ and $\text{Nd}:\text{YVO}_4$ crystals for UOT, because of their short response times.^{23,24} However, these crystals usually work in a narrow range of wavelengths. The third type of method uses a pixel array, i.e., a camera, to detect the tagged light.^{9,25–28} Since the AC signal for each speckle grain is individually measured before being added together, this parallel method increases the modulation depth by \sqrt{N} compared with that of the original single-pixel detector based method, and achieves a much larger etendue. To increase the sensitivity, heterodyne holography^{9,29} was developed to boost the signal strength by a strong reference beam and thus achieved shot noise limited sensitivity.

In previous heterodyne holography based UOT,⁹ a camera recorded the beat formed by the tagged light and a planar reference beam at a frame rate of $4\times$ the beat frequency. If averaging is not performed, four frames are required to reconstruct the amplitude of the tagged light. Thus, the time to obtain a UOT signal corresponding to one ultrasonic scanning position is limited by the low frame rates of conventional cameras (typically < 700 Hz with 300×300 pixels). Such a low speed makes this method inappropriate for *in vivo* applications since speckles decorrelate faster than 1 ms for thick living biological tissue (primarily due to blood

^{a)} Author to whom correspondence should be addressed. Electronic mail: lhwang@wustl.edu.

flow). Moreover, conventional cameras record both the AC signal and the DC background for each pixel. When the modulation depth is low, which is the case in UOT since the tagged photons are buried in a large background of untagged photons, only a few bits of a pixel value can be used to represent the useful AC signal, while most of the bits are wasted in representing the informationless DC background, resulting in a low efficiency in the use of bits. For example, even with a 16-bit analog-to-digital converter (ADC), no more than 3 bits of a pixel value can be used to represent the signal, when the modulation depth is lower than 10^{-4} for each pixel. A modulation depth as low as 10^{-4} is realistic when UOT targets 10 mm deep inside chicken breast tissue, using a 50 MHz ultrasonic transducer (UT) for high spatial resolution.³⁰ Besides the low bit efficiency, in previous methods all 16-bit data for each pixel, including both the small signal and the large background, were transferred to a computer. Since at least four frames of images must be transferred to calculate the UOT signal, the speed of data acquisition was severely limited by the heavy data transfer load.

In this work, we overcome the above drawbacks in heterodyne holography based UOT by using a lock-in camera^{31–33} (heliCam C3, Heliotis; 300×300 pixels, $40 \mu\text{m}$ pixel size), in which each pixel performs analog lock-in detection and outputs only the information of the AC signal, with a frame rate of up to 3800 frames per second to an on-chip memory. Specifically, the lock-in circuitry generates the in-phase ($S_I(r_i)$) and the quadrature ($S_Q(r_i)$) components of the AC signal for each pixel with a position of r_i , $i = 1, 2, \dots, 300 \times 300$, which are digitized by a 10-bit ADC. Then, using an on-chip field-programmable gate array (FPGA), the sought-after AC amplitude map $A(r_i)$ is calculated by $\sqrt{S_I^2(r_i) + S_Q^2(r_i)}$ and transferred to a computer. Since the information of only the AC signal, not that of the DC background, is digitized, the lock-in camera dramatically increases the bit efficiency by using all the bits to represent the signal, and it enables the use of inexpensive low-resolution ADCs. Moreover, compared with conventional camera based heterodyne holography that needs to record and transfer 4 frames, our approach is able to obtain the AC amplitude map after only a single frame of measurement taken within 0.3 ms. It also reduces the amount of data to transfer by transmitting only one frame of the calculated AC amplitude map, instead of four frames of raw images composed of both the signal and the background. Moreover, the data processing on the computer is much simpler and faster, since we need only to add up the AC amplitudes of all pixels to obtain the UOT signal S : $S = \sum_{i=1}^M A(r_i)$, where $M = 300 \times 300$.

Morgan *et al.* developed a prototype sensor aimed for lock-in camera based UOT,³² and Barjean *et al.* recently developed a Fourier transform method to improve the axial resolution of this approach.³⁴ However, both groups used custom-designed sensors with a relatively low etendue (64×64 and 24×24 pixels), and so far few experiments in the UOT field have been done with dynamic samples which more closely resemble living biological tissue. Here, using a commercially available high-resolution lock-in camera, we performed UOT inside a dynamic scattering medium with a high-frequency ultrasonic transducer. We were able to acquire an image of an absorptive object (AO) with a lateral

resolution of $153 \mu\text{m}$, even when the speckle correlation time was as short as $26 \mu\text{s}$, which is the shortest speckle correlation time ever reported in the UOT field.

The set-up for lock-in camera based UOT is schematically shown in Fig. 1. The output of a 532 nm continuous-wave laser (4 W, Verdi V5, Coherent) with a frequency of f_0 was split into a reference beam and a sample beam. After passing through acousto-optic modulators 1 (AOM1, AOM-505AF1, IntraAction) and 2 (AOM2, AOM-802AF1, IntraAction), the frequencies of the reference beam and the sample beam became $f_0 + f_{a1}$ and $f_0 + f_{a2}$, where $f_{a1} = 50 \text{ MHz}$ and $f_{a2} = 75 \text{ MHz}$ were the frequencies of the signals to drive the AOMs. The sample beam illuminated a dynamic scattering medium composed of two tissue-mimicking intralipid-gelatin phantoms³⁵ (IP1 and IP2). IP1 ($25 \text{ mm} \times 50 \text{ mm} \times 1.5 \text{ mm}$ along the x -, y -, and z -directions) and IP2 ($60 \text{ mm} \times 60 \text{ mm} \times 3 \text{ mm}$ along the x -, y -, and z -directions) had a reduced scattering coefficient (μ_s') of 1 mm^{-1} . An absorptive object (AO, $2 \text{ mm} \times 1.6 \text{ mm} \times 1 \text{ mm}$ along the x -, y -, and z -directions, absorption coefficient = 5.4 mm^{-1}) made of gelatin and black ink was attached to the surface of IP2, and the distance between the object and IP1 was 10 mm. During the integration time of the lock-in camera ($286 \mu\text{s}$, the minimum for our configuration), a 25 MHz spherically focused ultrasonic transducer (UT, V324-SM, Olympus), driven by a 200 V_{PP} sinusoidal wave with a frequency of $f_{a2} - f_{a1} - f_b = 24.93 \text{ MHz}$, emitted ultrasound along the x -direction to shift the frequency of a portion of the light passing through the ultrasonic focus to $f_0 + f_{a1} + f_b$, where $f_b = 70 \text{ kHz}$. These tagged photons beat with the reference beam at a frequency of f_b , and the lock-in camera was used to extract the AC amplitude of the beat (which was proportional to the amplitude of the tagged light) in each pixel with only a single frame of measurement. The reference beam and the sample beam (untagged light plus tagged light) on the camera had 24 mW and $53 \mu\text{W}$ of power, respectively. Since the shot noise was larger than the




FIG. 1. Schematic of the set-up for lock-in camera based UOT inside a dynamic scattering medium. AMP, power amplifier; AO, absorptive object; AOM, acousto-optic modulator; BB, beam block; BS, beamsplitter; FG, function generator; HWP, half-wave plate; IP, intralipid-gelatin phantom; L, lens; M, mirror; PBS, polarizing beamsplitter; R, reference beam; S, sample beam; TS, motorized translation stage; UT, ultrasonic transducer.

electronic noise of the camera and the technical noise of the laser (see [supplementary material](#)), we achieved shot noise limited sensitivity.⁹ The ultrasonic transducer and the center of the object shared the same z position, and the position of the lock-in camera was adjusted so that roughly each pixel detected one speckle grain.

By translating IP1 along the y -axis at different speeds with a motorized stage, we were able to control the correlation time of speckles on the lock-in camera.²² To calibrate the relationship between the phantom movement speed and the speckle correlation time, we placed a conventional camera (pco.edge 5.5, PCO-TECH; pixel size = $6.5 \mu\text{m}$) in the mirrored plane of the lock-in camera (see Fig. 1), to record movies of speckles when the phantom was moved at different speeds. A 6×6 pixel-binning was used to match the pixel sizes of the two cameras. Then, we calculated the correlation coefficients between the first and each of the ensuing frames of the recorded speckle patterns. By fitting the correlation coefficient versus time using^{22,36} $y(t) = a \exp(-2t^2/\tau_c^2) + b$, we obtained the speckle correlation time τ_c , defined as the time during which the correlation coefficient decreased to $1/e^2$ at a given phantom movement speed. For example, Fig. 2(a) shows the correlation coefficient as a function of time when the phantom was moved at 0.01 mm/s , from which $\tau_c = 59 \text{ ms}$ was determined. The relationship between the speckle correlation time τ_c and the phantom movement speed v is shown in Fig. 2(b). By fitting the experimental data with a theoretical model²² $\tau_c = d/v$, we obtained $\tau_c = 0.52/v$ [ms] (the unit of v is mm/s). Based on this equation, we were able to pre-set the speckle correlation time by controlling the phantom movement speed.

Figure 3 shows the UOT signal ($S = \sum_{i=1}^M A(r_i)$) and the root mean square (rms) UOT signal ($S_{\text{rms}} = M \sqrt{\sum_{i=1}^M A^2(r_i)/M} = M \sqrt{\sum_{i=1}^M [S_I^2(r_i) + S_Q^2(r_i)]/M}$) as a function of the speckle correlation time τ_c after self-normalization. When τ_c is longer than the integration time of the lock-in camera ($\tau_{\text{int}} = 286 \mu\text{s}$), the signal level is almost constant. In other words, speckle decorrelation does not affect the measurement, which is expected. When $\tau_c < \tau_{\text{int}}$, however, the UOT signal decreases with decreasing τ_c . Surprisingly, the decay of the UOT signal is rather slow. Even when $\tau_c = 26 \mu\text{s}$, which is $\tau_{\text{int}}/11$, the signal decreases only to around half of the signal obtained when $\tau_c = 100 \text{ ms}$.

To better understand the relationship between the UOT signal and the speckle correlation time, we developed the




FIG. 2. (a) The correlation coefficient between the speckle patterns as a function of time, when the phantom was moved at 0.01 mm/s . (b) The relationship between the speckle correlation time and the phantom movement speed. Error bars are not plotted due to their indiscernible lengths in the figure.




FIG. 3. Normalized UOT signal as a function of speckle correlation time. Error bars are not plotted due to their indiscernible lengths.

following analytical model. Let N denote the number of cycles integrated in the demodulation process of the lock-in camera. Then $N = \tau_{\text{int}}/\tau_0$, where τ_0 is the period for one cycle at the lock-in frequency f_b , i.e., $\tau_0 = 14.3 \mu\text{s}$. In our case, $N = 20$. Based on the principle of the lock-in camera, the in-phase signal S_I for each pixel can be expressed as

$$S_I(\tau_c) = \sum_{m=1}^N s_I(m\tau_0), \quad (1)$$

where $s_I(m\tau_0)$ is the in-phase signal for the m -th cycle in the beat. The autocorrelation function can be expressed as³⁶ $g^{(1)}(\tau) = \langle s_I^*(t)s_I(t+\tau) \rangle / s_{10}^2 = \exp[-(\tau/\tau_c)^2]$, where $\langle \cdot \rangle$ denotes ensemble averaging over all camera pixels, τ is a time delay, and s_{10}^2 is a constant. After squaring both sides of Eq. (1) and taking the ensemble average over all camera pixels, we obtained $\langle S_I^2(\tau_c) \rangle = N s_{10}^2 + 2 s_{10}^2 \sum_{n=1}^{N-1} (N-n) \exp[-(n\tau_0/\tau_c)^2]$. When $\tau_c \gg \tau_{\text{int}}$, $\exp[-(n\tau_0/\tau_c)^2] \approx 1$, so $\langle S_I^2(\tau_c) \rangle_{\tau_c \gg \tau_{\text{int}}} \approx N^2 s_{10}^2$. $\langle S_Q^2(\tau_c) \rangle$ can be derived similarly. Thus, the normalized rms UOT signal can be calculated by $\sqrt{\langle S_I^2(\tau_c) \rangle} / \sqrt{\langle S_I^2(\tau_c) \rangle_{\tau_c \gg \tau_{\text{int}}}}$ $= \sqrt{1/N + 2/N^2 \sum_{n=1}^{N-1} (N-n) \exp[-(n\tau_0/\tau_c)^2]}$, which is the analytical model used in Fig. 3. Although the theoretical




FIG. 4. One-dimensional images of the embedded object acquired under different conditions. The data points denote the experimental data, which are fitted to a theoretical model (denoted by the lines). The signals are normalized by the maximum signal in each condition.

predictions do not match the experimental data extremely well, the trends and the shapes are fairly similar.

To acquire an image of the absorptive object, we scanned the object along the y -axis and detected the UOT signal at each scanning position. During the whole process, IP1 was moved back and forth along the y -axis to cause the speckles on the lock-in camera to decorrelate. We took measurements only when the motorized stage reached a steady speed. Fig. 4 shows one-dimensional (1D) images of the object acquired when the speckle decorrelated at different rates. The dips in the images represent the object, which absorbed part of the tagged light. Although the signal-to-noise ratio is poorer when $\tau_c = 26 \mu\text{s}$, the image qualities for $\tau_c = 26 \mu\text{s}$ and $\tau_c > 50 \text{ s}$ (static phantom) are comparable in terms of image contrast, resolution, and the object's width and position, demonstrating that we were able to acquire an image of the object even when $\tau_c = 26 \mu\text{s}$. Fig. 4 also shows a direct transmission image in which the total transmitted sample light power measured by the conventional camera at each object scanning position is plotted. We could not observe the object in this image, since the resolution was too poor without doing UOT. To quantitatively determine the lateral image resolution r_{FWHM} of UOT (defined as the full-width at half maximum of the 1D point spread function), we fitted the experimental data with the following theoretical model:²² $y(x) = \text{cerf}[2\sqrt{\ln 2}(x-x_1)/r_{FWHM}] - \text{derf}[2\sqrt{\ln 2}(x-x_2)/r_{FWHM}]$, where c , d , x_1 , x_2 , and r_{FWHM} are fitting parameters, and $\text{erf}(x) = \frac{2}{\sqrt{\pi}} \int_0^x \exp(-t^2) dt$ is the error function. From the fitting, the lateral resolutions are $247 \mu\text{m}$, $286 \mu\text{m}$, and $267 \mu\text{m}$, when $\tau_c > 50 \text{ s}$, $= 1 \text{ ms}$, and $= 26 \mu\text{s}$. These values are 1.6–1.9 times as large as the measured FWHM ($= 150 \mu\text{m}$) of the 1D acoustic intensity profile, which is not uncommon and possibly due to the absorption of the tagged photons by the object when they are tagged in the vicinity of the object.^{12,37} If the spans between 25% and 75% of the contrasts are used,¹⁹ the lateral resolutions are $142 \mu\text{m}$, $163 \mu\text{m}$, and $153 \mu\text{m}$, when $\tau_c > 50 \text{ s}$, $= 1 \text{ ms}$, and $= 26 \mu\text{s}$. The object width can be obtained by $x_2 - x_1$, which gives 1.51 mm, 1.61 mm, and 1.62 mm when $\tau_c > 50 \text{ s}$, $= 1 \text{ ms}$, and $= 26 \mu\text{s}$. These values are very close to the true 1.6 mm width of the object.

In our set-up, there is a large gap ($= 10 \text{ mm}$) between the phantom IP1 and the transducer UT, due to the walls of a water tank and a phantom mount. Because of this distance, after passing through IP1, the light diffused to a big blob on the x - y plane where the transducer resides, even though IP1 was not thick compared with the samples used in previous studies. The modulation depth for each camera pixel was measured to be 0.12%, which is relatively low among the numbers reported in the literature.

Currently, even though the measurement is done within 0.3 ms, it takes at least 6 ms to transfer the data from the lock-in camera to the computer, limited by the low speed of the USB 2.0 interface ($\sim 250 \text{ Mb/s}$). If a camera link interface is used (data transfer rate $\sim 7024 \text{ Mb/s}$), the data transfer time can be reduced to $\sim 0.2 \text{ ms}$. Another way to lower the transfer time is to reduce the data load. Instead of transferring an AC amplitude map, it would be better to calculate the summed AC amplitude ($S = \sum_{i=1}^M A(r_i)$) using the on-chip FPGA,³⁴ and transfer only this number instead of one frame of 300×300 numbers. This simplification can greatly reduce the data load and thus the data transfer time.

The axial resolution was poor in our experiment, since a long burst of ultrasound was used to make sure the beat signal existed during the integration time of the lock-in camera ($286 \mu\text{s}$). To further improve the axial resolution, a frequency-swept method³⁸ or a Fourier transform method³⁴ can be employed.

To conclude, in this work, we used lock-in camera based UOT to acquire high-resolution images of an absorptive object inside a dynamic scattering medium. Our method is bit efficient. It can finish the measurement within $286 \mu\text{s}$, which is comparable with the typical speckle correlation time of living biological tissue ($100\text{--}1000 \mu\text{s}$), and it can tolerate speckle correlation time as short as $26 \mu\text{s}$. With these capabilities, our method holds promise for *in vivo* deep tissue non-invasive optical imaging.

See [supplementary material](#) for a noise analysis.

This work was sponsored in part by National Institutes of Health Grant Nos. DP1 EB016986 and R01 CA186567.

- ¹L. Wang and X. Zhao, *Appl. Opt.* **36**(28), 7277 (1997).
- ²D. S. Elson, R. Li, C. Dunsby, R. Eckersley, and M.-X. Tang, *Interface Focus* **1**(4), 632 (2011).
- ³S. G. Resink, A. C. Boccara, and W. Steenbergen, *J. Biomed. Opt.* **17**(4), 040901 (2012).
- ⁴L. Wang, S. L. Jacques, and X. Zhao, *Opt. Lett.* **20**(6), 629 (1995).
- ⁵S.-R. Kothapalli, S. Sakadzic, C. Kim, and L. V. Wang, *Opt. Lett.* **32**(16), 2351 (2007).
- ⁶L. V. Wang, *Phys. Rev. Lett.* **87**(4), 043903 (2001).
- ⁷W. Leutz and G. Maret, *Phys. B: Condens. Matter* **204**(1), 14 (1995).
- ⁸M. Kempe, M. Larionov, D. Zaslavsky, and A. Z. Genack, *J. Opt. Soc. Am. A* **14**(5), 1151 (1997).
- ⁹M. Gross, P. Goy, and M. Al-Koussa, *Opt. Lett.* **28**(24), 2482 (2003).
- ¹⁰S. Sakadzic and L. V. Wang, *Opt. Lett.* **29**(23), 2770 (2004).
- ¹¹S.-R. Kothapalli and L. V. Wang, *J. Biomed. Opt.* **13**(5), 054046 (2008).
- ¹²G. Rousseau, A. Blouin, and J.-P. Monchalain, *Opt. Lett.* **34**(21), 3445 (2009).
- ¹³Y. Li, P. Hemmer, C. Kim, H. Zhang, and L. V. Wang, *Opt. Express* **16**(19), 14862 (2008).
- ¹⁴Y. Li, H. Zhang, C. Kim, K. H. Wagner, P. Hemmer, and L. V. Wang, *Appl. Phys. Lett.* **93**(1), 011111 (2008).
- ¹⁵H. Zhang, M. Sabooni, L. Rippe, C. Kim, S. Kröll, L. V. Wang, and P. R. Hemmer, *Appl. Phys. Lett.* **100**(13), 131102 (2012).
- ¹⁶T. W. Murray, L. Sui, G. Maguluri, R. A. Roy, A. Nieva, F. Blonigen, and C. A. DiMarzio, *Opt. Lett.* **29**(21), 2509 (2004).
- ¹⁷F. Ramaz, B. Forget, M. Atlan, A. C. Boccara, M. Gross, P. Delaye, and G. Roosen, *Opt. Express* **12**(22), 5469 (2004).
- ¹⁸M. Gross, F. Ramaz, B. C. Forget, M. Atlan, A. C. Boccara, P. Delaye, and G. Roosen, *Opt. Express* **13**(18), 7097 (2005).
- ¹⁹P. Lai, X. Xu, and L. V. Wang, *J. Biomed. Opt.* **17**(6), 066006 (2012).
- ²⁰M. Gross, P. Goy, B. Forget, M. Atlan, F. Ramaz, A. Boccara, and A. Dunn, *Opt. Lett.* **30**(11), 1357 (2005).
- ²¹A. Lev and B. Sfez, *J. Opt. Soc. Am. A* **20**(12), 2347 (2003).
- ²²Y. Liu, P. Lai, C. Ma, X. Xu, A. A. Grabar, and L. V. Wang, *Nat. Commun.* **6**, 5904 (2015).
- ²³S. Farahi, G. Montemezzani, A. A. Grabar, J.-P. Huignard, and F. Ramaz, *Opt. Lett.* **35**(11), 1798 (2010).
- ²⁴B. Jayet, J. P. Huignard, and F. Ramaz, *Opt. Express* **22**(17), 20622 (2014).
- ²⁵S. Lévêque, A. C. Boccara, M. Lebec, and H. Saint-Jalmes, *Opt. Lett.* **24**(3), 181 (1999).
- ²⁶G. Yao, S. Jiao, and L. V. Wang, *Opt. Lett.* **25**(10), 734 (2000).
- ²⁷J. Li and L. V. Wang, *Appl. Opt.* **41**(10), 2079 (2002).
- ²⁸J. Li, G. Ku, and L. V. Wang, *Appl. Opt.* **41**(28), 6030 (2002).
- ²⁹F. Le Clerc, L. Collot, and M. Gross, *Opt. Lett.* **25**(10), 716 (2000).
- ³⁰M. Jang, H. Ruan, B. Judkewitz, and C. Yang, *Opt. Express* **22**(5), 5787 (2014).
- ³¹S. Bourquin, P. Seitz, and R. P. Salathé, *Opt. Lett.* **26**(8), 512 (2001).
- ³²S. P. Morgan, C. Li, B. R. Hayes-Gill, N. B. E. Sawyer, and C. Kongsavatsak, *Proc. SPIE* **6086**, 608615 (2006).

- ³³Y. Liu, C. Ma, Y. Shen, and L. V. Wang, *Opt. Lett.* **41**(7), 1321 (2016).
- ³⁴K. Barjean, K. Contreras, J.-B. Laudereau, É. Tinet, D. Etori, F. Ramaz, and J.-M. Tualle, *Opt. Lett.* **40**(5), 705 (2015).
- ³⁵P. Lai, X. Xu, and L. V. Wang, *J. Biomed. Opt.* **19**(3), 035002 (2014).
- ³⁶D. D. Duncan and S. J. Kirkpatrick, *J. Opt. Soc. Am. A* **25**(8), 2088 (2008).
- ³⁷G. Rousseau, A. Blouin, and J.-P. Monchalin, *Proc. SPIE* **7564**, 75642A (2010).
- ³⁸L. V. Wang and G. Ku, *Opt. Lett.* **23**(12), 975 (1998).